

**Iso-Analytical Limited*****Report of Analysis*****IA-R004 – <sup>13</sup>C Corn Flour Laboratory Standard**

This laboratory standard is intended to provide a sample of known isotope composition with a <sup>13</sup>C/<sup>12</sup>C isotope ratio stated in parts per thousand difference (‰) from the V-PDB (Pee Dee Belemnite) isotope ratio standard. This laboratory standard is not certified, but is provided to allow routine checking of the overall quality of measurements performed by continuous-flow isotope ratio mass spectrometry, and may be used as part of a quality control program. It is not intended for use as a substitute for calibration materials or inter-comparison materials distributed by NIST or IAEA.

*Analysis*

The <sup>13</sup>C/<sup>12</sup>C isotope ratio of the laboratory standard was measured by elemental analyser continuous-flow isotope ratio mass spectrometry using IAEA-CH-6 (ANU Sucrose) as the calibration material. The <sup>13</sup>C/<sup>12</sup>C isotope ratio in the laboratory standard was measured five times on three separate occasions.

*Isotope Abundance*

The laboratory standard IA-R004 is compared to V-PDB for the <sup>13</sup>C/<sup>12</sup>C isotope ratio. The isotope composition of the laboratory standard in ‰ relative to V-PDB is:

Laboratory Standard	$\delta^{13}\text{C}_{\text{V-PDB}} (\text{‰})$ $\delta_m \pm \sigma_1$
IA-R004	-10.99 ± 0.06

Note:  $\delta_m = \sum_{i=1}^n \delta_i/n$  ;  $\sigma_1 = \sqrt{\sum_{i=1}^n (\delta_m - \delta_i)^2/(n-1)}$  ; n = 15 for <sup>13</sup>C

April 23, 2003

Steven T. Brookes

Charles Belanger